

計量室ユニットのQ3器差再現性

新JIS20mm

サイト一技研(株)

指示ユニット1個、羽根車ユニット1個、および計量室ユニット9個を用い(調整器位置全て同じ)

6台の下ケースにて器差の再現性を試験

Q3流量: 4000l/h

水温: 20±1°C

2012/8/22

下ケース ↓	計量室ユニット								
	①	②	③	④	⑤	⑥	⑦	⑧	⑨
No.1 A社12B	-0.1	0.2	0.1	0.1	0.0	0.0	0.0	0.0	0.0
	-0.3	0.2	0.2	-0.1	0.0	0.1	0.1	-0.1	-0.3
	0.0	0.1	0.1	0.0	0.2	0.0	0.1	0.0	0.0
No.2 A社12B	-1.6	-1.4	-1.2	-1.3	-1.6	-1.2	-1.3	-1.5	-1.4
	-1.7	-1.4	-1.2	-1.6	-1.4	-1.3	-1.2	-1.1	-1.4
	-1.6	-1.6	-1.2	-1.3	-1.6	-1.2	-1.0	-1.5	-1.1
No.3 B社11B	-1.0	-1.0	-0.7	-0.7	-0.9	-0.8	-0.9	-1.0	-0.9
	-1.1	-1.0	-0.8	-0.9	-0.9	-0.8	-0.8	-1.0	-0.7
	-1.1	-1.0	-0.9	-0.8	-0.9	-0.8	-1.0	-0.9	-0.7
No.4 B社11B	-0.6	-0.4	-0.1	-0.3	-0.6	-0.3	-0.3	-0.5	-0.7
	-0.4	-0.5	-0.3	-0.2	-0.6	-0.4	-0.5	-0.7	-0.5
	-0.6	-0.2	0.0	-0.3	-0.7	-0.5	-0.5	-0.5	-0.4
No.5 C社08B	0.2	0.6	0.7	0.7	0.2	0.6	0.6	0.5	0.6
	0.1	0.5	0.7	0.8	0.3	0.7	0.7	0.6	0.5
	0.4	0.6	0.8	0.8	0.5	0.5	0.5	0.5	0.6
No.6 C社08B	-0.1	-0.1	0.0	0.1	-0.1	0.0	0.0	-0.1	0.0
	0.0	0.0	0.0	0.2	0.0	0.0	0.3	-0.2	-0.1
	0.0	0.0	0.2	0.0	0.0	0.0	0.1	-0.1	0.0